

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/19/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21



SGS Ruffly, Barrett WCSO #162 12-19-12