

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:40
INTERNAL STANDARD	VERIFIED	02:40
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44



SGS RJP Bartlett

WES0 #162 12-21-13