

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/02/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:47
INTERNAL STANDARD	VERIFIED	03:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:50



SGT Ralph J Bartlett weso #162 01-02-13