

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:35
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39



SG *Ralph Bartlett* *RJ* Waco #102 01-21-13