

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:12
INTERNAL STANDARD	VERIFIED	03:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:15



See Rpt / Bartlett

WCSD #162

01-30-13