

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/12/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:09
INTERNAL STANDARD	VERIFIED	01:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13



SGS Ralph J Bartlett

WESO #162

02-12-13