

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/18/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:17
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:20



SGT *Ralph Bartlett* Weso #162 02-18-13