

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/04/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:33
INTERNAL STANDARD	VERIFIED	02:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:36



SGS RALPH J BARTLETT WARSO #1062 03-04-13