

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/12/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:21 |
| INTERNAL STANDARD | VERIFIED | 01:21 |
| REFERENCE STANDARD TEMP | 33.97c | |
| REFERENCE STANDARD | 0.09 | 01:21 |
| BLANK TEST | 0.00 | 01:22 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 01:22 |
| BLANK TEST | 0.00 | 01:23 |
| REFERENCE STANDARD TEMP | 33.98c | |
| REFERENCE STANDARD | 0.09 | 01:23 |
| BLANK TEST | 0.00 | 01:24 |



SW Rep of Barry WCSO #162

03-12-13