

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/17/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:36
INTERNAL STANDARD	VERIFIED	17:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:36
BLANK TEST	0.00	17:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:38
BLANK TEST	0.00	17:39



SOB Ralph Bartlett #162 03-17-13