

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/09/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:42
INTERNAL STANDARD	VERIFIED	02:42
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:46



SGT *RJ Bartlett*

WASO #162

04-09-13