

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/23/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:59



505 *Ralph Bartlett* Waco #162 04-23-B