

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/08/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:11
INTERNAL STANDARD	VERIFIED	07:11
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	07:13
BLANK TEST	0.00	07:14



SGS

*Ralph Bartlett*

WCSO #162

05-08-13