

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/13/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 130560

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:12
INTERNAL STANDARD	VERIFIED	03:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:15
BLANK TEST	0.00	03:15



SGT RJA / T. Smith WSO #162 05-13-13