

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:29
INTERNAL STANDARD	VERIFIED	02:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:32



565 *RJ Bartlett*

QC 150 # 162 05-21-13