

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/27/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:13
INTERNAL STANDARD	VERIFIED	04:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:14
BLANK TEST	0.00	04:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:15
BLANK TEST	0.00	04:16



SG R J Bartlett WCSO #162 05-27-13