

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/24/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:11
INTERNAL STANDARD	VERIFIED	03:11
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:14



SGS *Ralph Bartlett* WESU#102 06-24-13