

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 07/02/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	02:59



565 *Ralph J Bartlett* Wase #162 07-02-13