

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/24/2013

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25

J. M. Lopez H290

