

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:27
INTERNAL STANDARD	VERIFIED	06:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:28
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	06:29
BLANK TEST	0.00	06:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:30
BLANK TEST	0.00	06:30



565 *Ralph J Bartlett*

WCSO # 162

07-30-13