

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/27/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:13
INTERNAL STANDARD	VERIFIED	02:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:13
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16



Set Rst / Injector

WCSO #162

08-27-13