

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/02/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:21
INTERNAL STANDARD	VERIFIED	04:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:22
BLANK TEST	0.00	04:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:23
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25



SGT Ralph Bartlett WACO #162

09-02-13