

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/16/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:52
INTERNAL STANDARD	VERIFIED	02:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56



SG- RJA/B... WC10 #10L 09-16-13