

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 09/24/2013

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:18
INTERNAL STANDARD	VERIFIED	02:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21



SW RJ Bartlett WSC #162 09-24-13