

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/30/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:59
INTERNAL STANDARD	VERIFIED	02:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03



565 Sgt J. Rutter W.C.S.O. #162 09/30/13