

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/08/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:40
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:43
BLANK TEST	0.00	23:44



SGT *Ralph Bartlett* 10/8/13

WCSO #162