

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/14/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24



SGS *RJ Bartlett* WCSO #162 10-14-13