

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/31/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:00
INTERNAL STANDARD	VERIFIED	22:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:01
BLANK TEST	0.00	22:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:02
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:03
BLANK TEST	0.00	22:03



505 Rpt of Bartlett w/sgo #162 12-31-13