

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 01/24/2014

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56



505 *Ralph Bartlett* WAGO #162 01-24-14