

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/28/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:07
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10



SGS Dept of Justice WCSO #162 01-28-14