

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 02/03/2014

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35



SGS By *R Bartlett* waso #102 02-03-14