

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/17/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:30
INTERNAL STANDARD	VERIFIED	01:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:32
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34



505 Bgl 1 Results were SHERIFF #162 02-17-14