

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 03/03/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57



SUS BY J. [unclear] WESB #102 03-03-14