

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 03/11/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:20



*sup by [signature] wesa #162 03-11-14*