

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/25/2014

Operator Name: KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00

[Handwritten Signature] #301 3/25/14

