

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/27/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:15
INTERNAL STANDARD	VERIFIED	04:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:15
BLANK TEST	0.00	04:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:17
BLANK TEST	0.00	04:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:18
BLANK TEST	0.00	04:18



SGS High Butter # 162