

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/24/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:16
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20



SGT Ralph J Bartlett STATION # 102 6/24/09