

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/02/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:57
INTERNAL STANDARD	VERIFIED	01:58
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01



*SOS Ralph Bartlett*

*weso #062*