

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/08/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:34
INTERNAL STANDARD	VERIFIED	00:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:36
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:37
BLANK TEST	0.00	00:38



SG Ralph J Bartlett #102 7/8/09