

SUPERVISOR TEST




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208

Date: 07/23/2009

Operator Name: KENNETH W SMITH  
 Permit #: 35604  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07



*KWSG* #301

07/23 /2009