

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/29/2009

Operator Name: KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:20
BLANK TEST	0.00	00:21



Kenneth W Smith 301

07/29/2009