

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/08/2009

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:43



*J.O. Lopez #299 8/8/09*