

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/19/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:47
INTERNAL STANDARD	VERIFIED	02:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:50



SGS Rugby Brewery WESO #102