

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/10/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:39
INTERNAL STANDARD	VERIFIED	01:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:42



SGT Ralph Bartlett WACO #162

9/10/09