

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/16/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:18
INTERNAL STANDARD	VERIFIED	02:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21



SGT Ralph J Bartlett 9/16/09

WCSO Sheriff #162