

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/24/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:27
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:30



505 *Ralph Bartlett*

9/24/09

WCSD SHEAF # 162