

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:22
INTERNAL STANDARD	VERIFIED	03:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:25
BLANK TEST	0.00	03:25



SGT *Ralph Bartlett*

WC SO # 162

10/17/09