

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/10/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:06
INTERNAL STANDARD	VERIFIED	04:06
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:06
BLANK TEST	0.00	04:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:09
BLANK TEST	0.00	04:09



SGS *Ralph Bartlett* 11/10/09 WOSO #162