

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/25/2009

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:08
INTERNAL STANDARD	VERIFIED	18:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:09
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:11

