

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/03/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:46 |
| INTERNAL STANDARD | VERIFIED | 00:46 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 00:46 |
| BLANK TEST | 0.00 | 00:47 |
| REFERENCE STANDARD TEMP | 33.97c | |
| REFERENCE STANDARD | 0.09 | 00:47 |
| BLANK TEST | 0.00 | 00:48 |
| REFERENCE STANDARD TEMP | 33.95c | |
| REFERENCE STANDARD | 0.09 | 00:49 |
| BLANK TEST | 0.00 | 00:49 |



SGA Ralph Bartlett

WCSD SHERIFF # 162

12/12/09